An Improved MRI Image Denosing Using Global Thresholding Technique

Manoj Gabhel¹, Rikhee Ram²

¹M.Tech, Dr. C.V. Raman Institute of Science and Technology, Bilaspur (C.G) India

²M.Tech, Dr. C.V. Raman Institute of Science and Technology, Bilaspur (C.G) India

Abstract: In medical image processing, image denoising plays very vital role in all through the diagnose. Medical images are usually corrupted by random noise in its acquisition and Transmission. Removing noise from the original signal is still a Bottleneck problem for researchers. The main objective of Image denoising techniques is necessary to remove such noises while retaining as much as possible the important signal features. There have been various denoising techniques; each has its own assumptions, advantages, and limitations. In this paper, a new thresholding technique using wavelet has been presented alongwith the standard thresholding techniques like soft and hard thresholding. Magnetic resonance (MR) images are routinely used for medical diagnosis. Denoising of these images to enhance their quality & Clinical parameter for an active area of research. This paper presents the wavelet-based thresholding scheme image denoising and noise suppression in MRI images. The performance of denoising scheme is evaluated in terms of PSNR, MSE and MAE.

Keywords: MRI, Thresholding, DWT, PSNR, MSE, MAE, Decomposition, Wavelet.

1. Introduction

Estimating a signal that is corrupted by additive noise has been of interest to many researchers for practical as well as theoretical reasons. The purpose is to recover the original signal from the corrupted or noisy data. The main aim of Image denoising techniques is to recover signal to be as close as possible to the original signal, while retaining its most important features (e.g. smoothness) and quality as much as possible. Magnetic Resonance imaging is a widely used medical imaging procedure because it is economical, comparatively safe, transferable, and adaptable. Though, one of its main shortcomings is the poor quality of images, which are affected by random noise. Traditional denoising schemes are based on linear methods, where the most common choice is the Wiener filtering. Recently, nonlinear methods, especially those based on wavelets have become increasingly popular [1].

One of the earliest papers in the field of wavelet-based denoising may be that of Weaver, et. al. [1]. In this pioneering work, they proposed a new method for filtering noise from MR (Magnetic Resonance) images based on the so-called hard-thresholding scheme. They showed that by using wavelet-thresholding, the noise could be significantly reduced without reducing the edge sharpness [2]. While Weaver, et al. demonstrated the advantages of the wavelet denoising scheme mainly based on experimental results, Donoho and Johnstone proved several important theoretical, results on wavelet thresholding, or wavelet shrinkage [3][4]. They showed that wavelet shrinkage has many excellent properties, such as near optimality in minimax sense, and a better rate of convergence [3][4]. DeVore and Lucier have also arrived at the wavelet thresholding concept, starting from their independent work on variational problems [5]. In particular, they were interested in finding an approximation ~ f to a given function f on a finite domain I that will balance the smoothes of the f & the closeness to the original function f. In order to find such f one tries to minimizes.

$$||f - g||_{L_2(I)}^2 + \lambda ||g||_{Y(I)}$$

over all g where Y is a space that measures the smoothness of the approximations g [5].

Besides wavelet-thresholding, many other approaches have been suggested as well. For example, wavelet-based denoising using Hidden Markov Trees [6], which was initially proposed by Crouse, et. al. has been quite successful, and it gave rise to a number of other HMT-based schemes. They tried to model the dependencies among adjacent wavelet coefficients using the HMT, and used the minimum mean-squared error (MMSE)-like estimators for suppressing the noise. Even though much work has been done in the field of wavelet thresholding, most of it was focused on the statistical modeling of wavelet coefficients for a certain class of signals (e.g. natural images), and the optimal choice of the threshold values. In this paper, we propose a new thresholding function that can take the place of the traditional thresholding functions, such as soft thresholding and hard-thresholding. We will demonstrate that the custom thresholding function outperforms the traditional ones, improving the denoised results significantly. Simulation results are given where appropriate, which show the advantage of the proposed scheme.

2. Discrete Wavelet Transform

Recently there has been significant investigations in medical imaging area using the wavelet transform as a tool for improving medical images from noisy data. Wavelet denoising attempts to remove the noise present in the signal while preserving the signal characteristics, regardless of its frequency content. As the discrete wavelet transform (DWT) corresponds to basis decomposition, it provides a non redundant and unique representation of the signal. Several properties of the wavelet transform, which make this representation attractive for denoising, are [7].

- **Multiresolution** image details of different sizes are analyzed at the appropriate resolution scales
- **Sparsity** the majority of the wavelet coefficients are small in magnitude.
- Edge detection large wavelet coefficients coincide with image edges.
- Edge clustering the edge coefficients within each sub band tend to form spatially connected clusters

During a two level of decomposition of an image using a scalar wavelet, the two-dimensional data is replaced with four blocks. These blocks correspond to the sub bands that represent either low pass filtering or high pass filtering in each direction. The procedure for wavelet decomposition consists of consecutive operations on rows and columns of the two-dimensional data. The wavelet transform first performs one step of the transform on all rows. This process yields a matrix where the left side contains down sampled low pass coefficients of each row, and the right side contains the high pass coefficients. Next, one step of decomposition is applied to all columns; this results in four types of coefficients, HH, HL, LH and LL.



Figure 1: Two-Level Image decomposition by using DWT

3. Introduction to Denosing

De-noising plays a very important role in the field of the medical image pre-processing. It is often done before the image data is to be analyzed. Denoising is mainly used to remove the noise that is present and retains the significant information, regardless of the frequency contents of the signal. It is entirely different content and retains low frequency content. De-noising has to be performed to recover the useful information. In this process much attention is kept on, how well the edges are preserved and how much of the noise granularity has been removed [8] [9].

The main purpose of an image-denoising algorithm is to eliminate the unwanted noise level while preserving the important features of an image. In wavelet domain, the noise is uniformly spread throughout the coefficients while mostly the image information is concentrated in the few largest coefficients. The most important way of distinguishing information from noise in the wavelet domain consists of thresholding the wavelet coefficients. Mainly hard and soft thresholdings techniques are performed.

4. Methodology

The reduction of noise present in images is an important aspect of image processing. Denoising is a procedure to recover a signal that has been corrupted by noise. After discrete wavelet decomposition the resulting coefficients can be modified to eliminate undesirable signal components. To implement wavelet thresholding a wavelet shrinkage method for de-noising the image has been verified. The proposed algorithm to be used is summarized in Algorithm 1 and it consists of the following steps

Algorithm 1: Wavelet image de-noising

- Choice of a wavelet (e.g. Haar, symmlet, etc) and number of levels or scales for the decomposition. Computation of the forward wavelet transform of the noisy image.
- Estimation of a threshold.
- Choice of a shrinkage rule and application of the threshold to the detail coefficients
- Application of the inverse transform (wavelet reconstruction) using the modified (threshold) coefficients.

For better and easy understanding, a complete flowchart of the discussed methodology has been shown above. The main algorithm, followed in order to fulfill the aim of this thesis, is as follows:

Step 1:

Read Simulated and original standard image (MRI.jpg, BRAIN.tif).

Step 2:

Resize the loaded image to a standard size of 256×256 . The images taken for test have a lot of variation in their sizes and hence cannot be compared on the same basis. For large sized images, such as 512×512 , the computation time for denoising is found to be more. And if the image size is taken smaller than 256×256 , then the useful data is liable to get lost.

International Journal of Science and Research (IJSR) ISSN (Online): 2319-7064 Impact Factor (2012): 3.358

5. Flow Diagram



Figure 2: Flowcharts for Image Denoising Algorithm Using

Wavelet Transform

Step 3:

Random Noise is added to the standard test images

Step 4:

Make the noisy image to undergo wavelet transform i.e. DWT.

Step 5:

Now select the desired Wavelet & Level. After that noisy image is decomposed into approximation and detail coefficients using wavelet transform.

Step 6:

Select the desired thresholding technique (Global, Level-Dependent & Optimal).After the decomposed image coefficients are threshold using the above mentioned three thresholding technique, the denoised image is reconstructed using inverse wavelet transforms- IDWT.

6. Thresholding Technique

Thresholding is the simplest method of image denoising .In this from a gray scale image, thresholding can be used to create binary image. Thresholding is used to segment an image by setting all pixels whose intensity values are above a threshold to a foreground value and all the remaining pixels to a background value. Thresholding is mainly divided into two categories:

A. Hard Thresholding : Hard threshold is a "keep or kill" procedure and is more intuitively appealing. The transfer function of the Hard thresholding is shown in the figure.

Hard thresholding may seem to be natural. Sometimes pure noise coefficients may pass the hard threshold and this thresholding method is mainly used in medical image processing.[10][11].



B. **Soft Thresholding :** Soft threshold shrinks coefficients above the threshold in absolute value. The false structures in hard thresholding can be overcomed by soft thresholding. Now a days, wavelet based denoising methods have received a greater attention. Important features are characterized by large wavelet coefficient across scales in most of the timer scales.[10]



Fig 4: Original and Soft thresholded signal

7. Selection of Thresholding Technique

In the application of De-noising, the threshold level parameter T plays an essential role.Values too small cannot effectively get rid of noise component, while values too large will eliminate useful signal components. There are a variety of ways to determine the threshold value T as we will discuss in this section Depending on whether or not the threshold value T changes across wavelet scales and spatial locations, the thresholding can be:

1. Global Threshold: The global threshold method derived by Donoho is given by Eq. (2) has a universal threshold:

$$\lambda = \sigma \sqrt{2 \log(N)}$$

Where N is the size of the coefficient arrays and σ^2 is the

Volume 3 Issue 9, September 2014

Paper ID: SEP14626

www.ijsr.net Licensed Under Creative Commons Attribution CC BY

2297

2)

noise variance of the signal samples.

(PSNR).

2. Level Dependent Threshold: Level dependent thresholding method is done by using Eq. (3).

Estimation of the noise standard deviation σ_k is done by using the robust median estimator in the highest sub-band of the wavelet transform

$$\lambda_k = \sigma_k \sqrt{2\log(N)}_{(3)}$$

Where the scaled MAD noise estimator is computed by

$$\sigma_k = \frac{MAD_k}{0.6745} = \frac{(median(|\omega_i|))_k}{0.6745}$$
(4)

Where MAD is the median absolute deviation of the magnitudes of all the coefficients at the finest decomposition scale and ω_i are the coefficients for each given sub-band, the factor 0.6745 in the denominator rescales the numerator so that σ_k is also a suitable estimator. The threshold estimation method is repeated for each sub-band separately, because the sub-bands exhibit significantly different characteristics.

3. Optimal Threshold Estimation: Estimate the mean square error function to that compute the error of the output to minimize the function, the minimum MSE serves as a solution to the optimal threshold. A function of the threshold value which is minimized is defined in Eq. (5).

$$G(\lambda) = MSE(\lambda) = \frac{1}{N} ||y - y_{\lambda}||^{2}_{(5)}$$

If y_{λ} is the output of the threshold algorithm with a threshold value λ and y is the vector of the clean signal, the remaining noise on this result equals $e_{\lambda} = y_{\lambda} - y$. As the notation indicates, the MSE is a function of the threshold value λ . Find the optimal value of λ that minimizes MSE (λ) and the convergence of the algorithm.

Step 7:

Then three parameters, PSNR (peak signal to noise ratio), MAE (mean absolute error) and MSE (mean square error) are calculated for all the standard images with their noisy and denoised counterparts, respectively. Hence, we get a good amount of comparison between the noisy and denoised images keeping the set standard image intact.

Step 8:

A usual way to de-noise is to find a processed image such that it minimizes mean square error MSE, MAE and increases the value of the PSNR. Hence depending upon the values of above three parameter, we conclude that which wavelet & Thresholding technique gives best denoised result.

8. Performance Evaluation

To get the measure of the wavelet performance, the experimental results are evaluated according to three error criteria namely, the mean square error (MSE), the mean absolute error (MAE) and the peak signal to noise ratio

1. *Peak signal to Noise Ratio (PSNR)*: PSNR measures the ratio between the maximum possible power of a signal and the power of corrupting noise that affects the quality and reliability of its representation. It defines the purity of the output signal. PSNR is calculated as follows:

$$PSNR = 10 \cdot \log_{10} \left(\frac{MAX_{I}^{2}}{MSE} \right)$$
$$= 20 \cdot \log_{10} \left(\frac{MAX_{I}}{\sqrt{MSE}} \right)_{(6)}$$

Where, MSE = Mean Squared Error, MAX_I is the maximum possible pixel value of the image

Mean Squared Error (MSE): Mean Square Error (MSE) function is commonly used because it has a simple mathematical structure that is easy to compute and it is differentiable implying that a minimum can be sought.

The MSE is the difference between the original image and the denoised image. Given by

$$MSE = \frac{1}{MN} \sum_{m=0}^{M-1} \sum_{n=0}^{N-1} \left(y(m,n) - \tilde{y}(m,n) \right)^2$$
(7)

3. *Mean of absolute error (MAE)* : Another criterion measure include: Mean of absolute error (MAE) which is given by

$$MAE = \frac{1}{MN} \sum_{m=0}^{M-1} \sum_{n=0}^{N-1} |y(m,n) - \hat{y}(m,n)|$$
(8)

The goal of de-noising is to find an estimate image such that MAE is minimum.

9. Results

For our test experiments we have considered an additive noise with a uniform distribution which has been used to corrupt our simulated and real MR test image objects. Artificially adding noise to an image allows us to test and assess the performance of various wavelet functions.

We used MATLAB to implement the de-noising algorithm. MATLAB has a wavelet toolbox and functions which are very convenient to do the DWT. A usual way to de-noise is to find a processed image such that it minimizes mean square error MSE, MAE and increases the value of the PSNR.

We have done simulations with uniform random noise added to the MR image. An example of a noisy magnetic resonance image (MRI) which consists of 256X256 pixels is shown in Fig. 5. As can be seen in the background the image has been uniformly corrupted with additive noise. The de-noising techniques discussed in the previous section are applied to the noisy MR image to test the efficiency of the different threshold methods.

International Journal of Science and Research (IJSR) ISSN (Online): 2319-7064 Impact Factor (2012): 3.358



Figure 5 (a): Original image and (b) Noisy image

For comparison of the five different wavelet functions, the quantitative de-noising results of the MRI images obtained by using global thresholding are shown in Table I and II respectively. The MSE, MAE, PSNR error criteria are the ones which have been used to assess the performance of the wavelet functions. Their numerical results are summarized in the tables.

Table 1: Qualitative analysis (MRI image) – GobalThresholding For Level 1

	Ŭ		
Type of Wavelet	Level 1		
	MSE	MAE	PSNR(db)
Haar	0.0089	0.0748	21.4544
db2	0.0082	0.0716	21.1694
db4	0.0082	0.0722	21.5369
sym2	0.0084	0.0722	21.2592
sym4	0.0084	0.0725	21.9326
bior1.1	0.0093	0.0761	21.3245
bior 1.3	0.0095	0.0770	20.7880

It is clear from the table I, for Global Thresholding technique; sym4 gives best result for level-1. Its gives higher PSNR & lower MSE & MAE value.

It is clear from the table II, for Global Thresholding technique; db4 performs well for level-2. Its gives higher PSNR & lower MSE & MAE value.

Table 2 : Qualitative analysis (MRI image) – GobalThresholding For Level 2

	<u> </u>		
Tupo of Wavalat	Level 2		
Type of wavelet	MSE	MAE	PSNR(db)
Haar	0.0130	0.0894	19.2090
db2	0.0117	0.0850	19.8857
db4	0.0119	0.0861	20.3062
sym2	0.0120	0.0867	20.0687
sym4	0.0116	0.0852	20.0149
bior1.1	0.0128	0.0886	18.7935
bior 1.3	0.0129	0.0888	19.3273

10. Conclusion and Future Scope

The de-noising process consists of decomposing the image, thresholding the detail coefficients, and reconstructing the image. The decomposition procedure of the de-noising example is accomplished by using the DWT. Wavelet thresholding is an effective way of de-noising as shown by the experimental results obtained with the use of different types of wavelets. Thresholding methods implemented comprised of the level (sub- band) thresholding and optimal thresholding. More levels of decomposition can be performed; the more the levels chosen to decompose an image, the more detail coefficients we get. But for denoising the noisy MR data sets, two-level decomposition provided sufficient noise reduction

In this paper we have presented the generalization of the DWT method for the 2-D case. The resulting algorithms have been used for the processing of noisy MR image. Experimental results have shown that despite the simplicity of the proposed de-noised algorithm it yields significantly better results both in terms of visual quality and mean square error values. Considering the simplicity of the proposed method, we believe these results are very encouraging for other forms of de-noising. The Biorthogonal wavelet (bior1.1) & Biorthogonal wavelet (bior1.3) gave the best results compared to other wavelets for both Simulated & MRI image respectively. Optimal thresholding gives better denoised result among the three thresholding technique.

As the future perspective can be seen, the mentioned methods can be implemented that to look how it can be used on different images. With different spatial resolution, different behaviours of same image would be quite interesting. Since selection of the right denoising procedure plays a major role, it is important to experiment and compare the methods

Reference

- [1] Byung Jun Yoon and P.P. Vaidyanathan, "Wavelet based denosing by customized thresholding", Work suppored in part by the ONR great N00014-99-1-1002, USA.
- [2] Weaver, J.B., Yansun, X Heavely, D. M. Jr., and Cromwell, L.D., "Filtering noise from images with wavelet transforms", Magnetic Resonance in Medicine, 24 (1991), 288-295.
- [3] Donoho, D. L. and Johnstone I. M., "Ideal spatial adaptation via wavelet shrinkage", Biometrika, vol. 81, pp. 425- 455, 1994.
- [4] Donoho, D. L., "De-noising by soft-thresholding", IEEE Trans. Info. Theory, vol. 41, pp. 613-627, May 1995.
- [5] DeVore, R. A. and Lucier, B. J., "Fast wavelet techniques for near-optimal image processing", IEEE Military Communications Conference Record, San Diego, Oct. 11-14, 1992, m vol. 3, pp. 1129-1135.
- [6] Crouse, M. S., Nowak, R. D., Baraniuk, R. G., "Waveletbased signal processing using hidden Markov models", IEEE Trans. Signal Proc., vol. 46, pp. 886-902 Apr. 1998.
- [7] S.Sudha, G.R.Suresh and R.Sukanesh, "Speckle Noise Reduction in Ultrasound Images by Wavelet Thresholding based on Weighted Variance", International Journal of Computer Theory and Engineering, Vol. 1, No. 1, April 2009 1793-8201.
- [8] S.Singh, Manish Kumar, "Denoising Of Natural Images Using The Wavelet Transform" (2010). Master's Theses.
- [9] S. G. Chang, B. Yu, and M. Vetterli, 2000 "Adaptive wavelet thresholding For image denoising and compression," IEEE Trans. Image Processing.,vol. 9, no. 9, pp. 1135–1151, sep.

Volume 3 Issue 9, September 2014

<u>www.ijsr.net</u>

- [10] Cajo J.F. ter Braak "Bayesian sigmoid shrinkage with improper variance priors and an application to wavelet denoising", Computational Statistics & Data Analysis vol. 51, pp. 1232 – 1242 / 2006.
- [11] C.O.S Sorzano, E.Ortiz, M.Lopez, J.Rodrigo "Improved bayesian image denoising based on wavelets with applications to electron microscopy/Pattern Recognition", vol. 39, pp.205 – 1213/2006.

Author Profile

Manoj Gabhel is currently pursuing his M.Tech (2012-14) from the department of Electronic & Communication Engineering Specialization in Digital Communication at Dr C.V.R.U. Bilaspur (C.G). He had done B.E in ET&T Engg. From K.I.T Raigarh C.G. in 2008 and PG Diploma in Industrial Automation System Design From NIELIT Calicut Kerala India in 2009. His Area of Interest includes Communication System. He has one year experience as a Lecturer in Government Engineering College Jagdalpur C.G. India.

Rikhee Ram, received the B.E. degree in Electronics and communication Engineering from Govt. Engineering College, affiliated to Guru Ghasidas University (presently known as Guru Ghasidas Vishwavidyalaya), Bilaspur (C.G.), India in 2007. Presently he receives M. Tech. in Digital Communication Engineering from Dr. C. V. Raman University, Kota, Bilaspur (C.G.), India. His area of interest in Signal Processing, Communication.