

Textile Capacitive Sensor Arrays with Hidden Markove Model and Dynamic Time Warping

Shubhangi S. Marudkar¹, Dr. Prashant R. Deshmukh²

¹Dept. of Computer Science and Engineering, Sipna College of Engineering and Technology, Amravati, India

²Professor, Dept. of Computer Science and Engineering, Sipna College of Engineering and Technology, Amravati, India

Abstract: Physical disabilities such as lower limb loss, manual dexterity, disable to co-ordinate with different organs of the body are very common sequel of brain injury, strokes, spinal cord injury and degenerative diseases such as affecting tissues or organs. Existing assistive technology solutions for such patients provide access as user input devices such technologies are intrusive and expensive, and require physical contact that can have hazardous effects such as skin friction injury for patients who have reduced their skin sensitivity. To overcome this problem, in this paper, we proposed the design and implementation of non-contact proximity gesture recognition system using textile capacitive sensor arrays. The Textile capacitive sensors can be easily integrated into items of clothing, bed sheets. This sensor is lightweight and flexible. To build this algorithm two techniques are used these are Hidden Markove Model and Dynamic Time Warping. These Techniques are used to convert raw capacitance values to alphanumeric gestures. This system is personalized to the user; this also allows the personalized selection of gesture and recognize gestures with varying complexity with less user training.

Keywords: Manual dexterity, Degenerative diseases

1. Introduction

Millions of individuals are hospitalized every year because of strokes, brain injuries and spinal cord injuries [1], [2], [3]. TRAUMATIC BRAIN INJURY (TBI) is an important global public health problem. About 10 million TBI patients are serious and results to death or hospitalization occur annually. The patients who live with TBI are almost the disabled persons who cannot do their regular work properly [2]. Severe impairments such as paralysis, paresis, weakness and limited range of motion are common sequelae resulting from these injuries and require extensive rehabilitation. These patients need assistive care devices for environmental control. For instance, assistive devices such as joysticks and switches, [4] are used for navigating wheelchairs, making calls to nurses in a hospital setting, and controlling appliances in the home. Such devices can help individuals maximize independence and provide methods to supplement direct care [5], [6]. Gesture recognition-based environmental control systems are a class of assistive devices capable of allowing individuals with mobility impairments a greater control over their environment. Several techniques such as the use of inertial sensors, vision systems, and tracking can be used to capture body gestures [7], [8], [9], [10], [11]. Sensors for gesture recognition are intrusive, bulky, and expensive [12]. Eye tracking systems, for instance, necessitate the use of mounted cameras while EEG electrode-based systems can cause skin irritation and abrasion, conditions that can have a deleterious effect if unnoticed due to diminished sensation in the extremities for these individuals. In reality, individuals perform gestures in many different ways. In fact, the same user may perform a gesture differently depending on the time of the day, medication, and fatigue. The caregivers assistive care systems should require absolute minimal setup and maintenance [1]. To Overcome all the problems we have designed, implemented and evaluated a system that uses wearable sensors built from textile capacitive sensor arrays (CSA) and a wrist worn accelerometer.[1] Our sensors work

on the principle of change in capacitance when there is movement in the proximity of the fabric capacitor plates. These plates can be used in the clothing fabrics and can be embedded into bed sheets and pillow covers [1].




Figure 1 demonstrates a 4 × 3 capacitive sensor array sewn into the denim fabric using conductive wires. It also demonstrates the accelerometer wristband used to find the orientation of the capacitor sensor array with respect to the hand. The data from the Sensors is analyzed using our custom-designed wireless module which uses capacitance measurement ICs, an MSP430 micro-controller, and Bluetooth wireless module.

Fig. 1 illustrates a prototype system built using the capacitive plates and conductive threads sewn into the fabric and the wrist band with an accelerometer. These sensors are proximity sensors and obviate the need for touch-based gesture recognition. An array of these sensors can be effectively used to track body positions and movements that can be consequently used to recognize gestures. We have designed and implemented a signal processing system that performs hand position tracking using the CSA, and converts a time series of hand positions into alphanumeric gestures using a Hidden Markov Model (HMM) and a model based on Dynamic Time Warping (DTW). This system learns how the gestures are performed by the user, and can adapt to different sensor configurations and placements with minimal training. The gestures are recognized irrespective of changes in orientation and placement of the sensors.

2. Related Work

The spotting task of gesture recognition is much more challenging, since it is difficult to spot human motion events from number of sources. Another challenge is the system has to deal with the motion events to be spotted may only occur in a continuous data stream, while at the same time embedded into other, These movements are difficult to model due to their complexity and unpredictability. The study of gesture recognition has been continued over years and many approaches have been proposed till now. These approaches are categorized in two ways, first is gesture recognition, requiring external infrastructure and the second is gesture recognition focusing on wearable instrumentation. The first category is the vision-based motion recognition in this multiple cameras can be used for gesture recognition. The second category is started working recently and over the few period of time this has gained much attention[7]. This system which we are using in this paper is based on the second category also it is based on the previous work on capacitive sensing, gesture recognition systems and signal processing algorithms.

A. Capacitive sensing technology

Capacitive sensing technology is very popular in industrial automotive and healthcare applications[13]. It has also been applied to positioning [15], humidity sensing[17], ambient sensing[19]. This technology used to replace optical detection methods and mechanical designs for the applications like gesture detection and material analysis. The main advantage of the capacitive sensing over the other gesture detection approaches are that it can sense different kinds of materials such as skin, plastic, metal etc. it is contactless and wear free, it also has the ability to sense large distance with small sensor sizes, it has low cost and low power solution. Capacitors as proximity sensors have applications to robotics, industrial monitoring and healthcare systems[13]. We present an end-to-end gesture recognition system including a 2D array and results from 2D tracking and gesture classification using a variable capacitor array[1]. Products like Microchip's PIC [24] allow for 3D gesture tracking using rigid capacitor arrays. The non-rigidity property of our sensor allows for sewing them into clothing and into the environment such as bed sheets, pillow covers, and wheelchair pads. Additionally, the use of capacitive sensors (CS) allows to work collaboratively to reduce environmental noise and capture motion attributes like speed and direction of motion[1].

B. Gesture Recognition System

Gesture recognition system interprets the human motion by computing device mathematically. Gesture recognition with facial recognition, voice recognition, lip movement and eye tracking are the components of which developers refer as perpetual user interface (PUI). The goal of PUI is to enhance the efficiency and ease the use of the underlying logical design of a stored program[1]. Most of the time the gestures are used for input commands. Recognizing gestures as input allows computers to be more accessible for the physically-impaired and makes interaction more natural in virtual environment. Hand and body gestures can be amplified by a controller containing accelerometers and gyroscopes to

sense tilting, rotation and acceleration of movement or the computing device can be outfitted with a camera so that software in the device can recognize and interpret specific gestures. A wave of the hand, for instance, might terminate the program. The increase in adoption of home automation technologies has spawned the new platforms for control based on gestures. for instance, allow environmental control without physical contact with a controlling device. Home automation techniques have been adapted to enable users with limited mobility to have greater autonomy over environmental control. These platforms include voice-activated systems [17], head-tracking [18], EOG based eye-tracking and inertial sensors [19]. In this signal processing algorithm adapts to the user and changes in the configuration of the sensors with minimal training. Additionally, unlike systems like cameras, this system is effortful and can be built into clothing and embedded into the environment such as pillows and bed sheets.

C. Signal Processing for Gesture Recognition

There is a large body of work on applying signal processing techniques to gesture recognition. Learning techniques such as Hidden Markov Models [20], decision trees [21], and Bayesian inference [22] have been applied to converting data from sensors such as accelerometers to movement activities and gestures. Learning approaches such as HMMs have also been applied to applications such as handwriting recognition [23]. We adopt two classification approaches, Hidden Markov Models and Dynamic Time Warping to our system to convert data from a capacitive sensor array into alphanumeric gestures.

3. Textile Capacitor Sensor Array For Gesture Recognition

The gesture recognition sensor consists of textile based electrodes which built into clothing or into the environment. Outdated users often lack of sensation in hands for them it is critical for regulating the contact, the repeated contact can be damaging to the skin to various degrees. Depending on the surface underlying. We also can create a touch-based swipe interface, in this the characters can be drawn directly on the fabric surface. The problem with this interface is that this can be develop skin friction injuries.




Figure 2: The gestures evaluated in this paper are based on EdgeWrite [26] gestures, for which users are taught to envision motions through corner areas in specific orders to encourage distinct gestures. Even so, robust classification is

required. This figure depicts two situations, from three gestures within our dataset, where a state machine approach does not suffice for classification.

As fig 2. Shows the user do not always perform the exact gestures as required. Therefore we have instructed the individuals using EdgeWrite gestures set to perform distinct gestures, in order to create a robust interface so that users does not require to direct full visual attention to the interface, but still we have need the more accurate and sophisticated gesture recognition algorithm than EdgeWrite. The figure shows two situations. In this EdgeWrite state machine approach will fail due to omission of states or adding extra states. Then we presents the touch less interface, where gestures can freely performed in the proximity of the sensor. Because of this we can perform gestures with frictionless and touch less. The signals detected at the proximity have lower power and higher variance than the touch based interfaces. In the touch interface the signals are easily thresholded to create binary detection in each point of the array. Where as in touch less interfaces binary thresholding is not reliable[1]. Capacitive sensors work on the principle of change in Capacitance the electric fields between the plates of capacitor makes them highly versatile. Accelerometers and gyroscopes measure movement of the body to which they are attached. capacitive sensors can sense movement of remote bodies.




Figure 3 (a)




Figure 3 (b)

Figure 3: (a) Equivalent electrical circuit when a capacitive plate is placed on the leg and the user performs gestures using his hand. The body is capacitively coupled to the sensor ground using a sensor ground plane
 Fig 3(b) Longitudinal cross sectional view of the sensor.

Fig 3(a) illustrates the principle using an example where an array of capacitive plates is worn by a user and gestures are performed by moving the hand in the vicinity of the array without touch. The body of the patient is capacitively coupled to the ground of the sensor system. When the hand is moved close to the capacitor plates, the capacitance C_b increases. Inversely, the value of capacitance C_b can be used

to localize the hand with respect to the plates. The range of our capacitive plates is sufficient to prevent accidental touch and skin abrasion and can be adapted as the range is variable with plate size and shape. In Fig 3(b) The longitudinal cross-section of our designed sensor is illustrated The top layer is a network of capacitive sensor plates connected via conductive threads. An AC shield plane minimizes parasitic capacitance and noise coupling. The ground plane capacitively couples the human body to the ground of the sensor and provides a common reference for the capacitance measurements.

4. Classification Algorithms for Gesture

There are many algorithms for gesture recognition. For instance, Optical Character Recognition (OCR) algorithms This is adapted for recognizing characters from the tracked hand positions. OCR, however, views the set of positions as an unordered set and does not average the temporal variations of the tracked position and direction of motion. Using both the temporal variation in the data and the direction of motion is critical in our application where the same gesture can be performed very differently at a time. In order to take advantage of the temporal ordering of the data, we chose to evaluate our system by adapting two techniques: Hidden Markov Models and Dynamic Time Warping. HMM is a statistical learning technique that performs well when there is large variations in the way the gestures are performed but requires substantial training and DTW is a time series based technique which requires minimal training. These two algorithms help us evaluate trade-offs in training, efficiency and accuracy of gesture recognition across multiple subjects.

A Hidden Markov Model

The Hidden Markov Models are used in almost all current speech recognition systems, in most of the applications such as computational molecular biology, in data compression and in other areas of Artificial Intelligence. A Hidden Markov Model is defined as a stochastic finite state machine. The algorithm is illustrated in Algorithm 1 Formally, it is defined by 5-tuple $\Omega=(S,\Sigma,\Pi,\delta,\lambda)$ where S represents a set of hidden states that are not directly observable. Transitions between the states is denoted by a transition probability matrix, δ . Π is the set of initial probabilities corresponding to the states in S . Every state has a set of possible emissions Σ and continuous probabilities λ for these emissions. The emissions can be observed by giving some information about the most similar underlying hidden state sequence which led to a particular sequence of observations. The HMM model assumes that the underlying physical process is Markovian, that means any prediction of future behavior can be optimally calculated by knowing the present state without any history. We map such a Hidden Markov Model to our problem of decoding a gesture based on a time series of hand position values. This definition of states constraints the set of gestures for which the Markovian assumption is valid. Our evaluation, however, demonstrates that this state definition works well across a large set of gestures.

Algorithm1. HMM Feed Forward(O, Φ, Σ, π, δ, λ, k)

Input: Observations (O): $\{(x_1, y_1), \dots, (x_n, y_n)\}$ for the gesture, Hidden states (S): $\{1, \dots, k\}$, Initial Probabilities: $\{\pi_1, \dots, \pi_k\}$, Transition Probabilities: $\delta(i, j), (1 \leq i, j \leq k)$, Emission probability distribution: $\lambda_1, \dots, \lambda_k$ (Gaussian (μ_k, σ) for every state $k \in [1, \dots, \text{len}(S)]$)
Output: p (Model Probability given the Observations)
for i: =1 **to** len(O) **do**
for k: =1 **to** len(S) **do**
 Emission = $(1/(\text{distance}((x_i, y_i), \mu_k)/\sigma)^2)+3$
for j: =1 **to** len(S) **do**
 $p(i, k) = p(i, k) + p(i - 1, j) \times \text{Emission}$
end for
end for
end for
return $\sum_k p(|O| - 1, k)$
Asymptotic Running time complexity: $O(|O| \times k^2)$

The transition probability matrix, δ is determined during a training phase when the subject performs a set of gestures. While calculating the probability of transitions between states, we make the following key adaptation to fit our problem domain. We have found that if the sampling frequency of our sensors is large compared to the transition rate, the transition probability matrix begins to resemble an identity matrix, and the importance of transitions between states is diminished.

The emission probabilities λ is calculated according to a Gaussian function. Each state is assigned a center position. The distance (D) from the estimated position and the center is used as the input for the Gaussian. The Gaussian width (σ) is approximately 1/3 of the center-to-center plate distance. This width is determined and used throughout the experiments. For computational efficiency, we use the following approximation $((D/\sigma)^2 + 3)^{-1}$ for the Gaussian. Applying the above modifications to the HMM, we generate a model per gesture. The gesture to be classified is passed through each of these models to create a vector of posterior probabilities relating to each Underlying model. The model with the highest probability is selected as the classification for the gesture.

B. Dynamic Time Warping

Hidden Markov model builds a statistical model for the gestures. It also accurately decodes gestures which have variations. This model must be trained on a large set of gestures. The Dynamic Time Warping is an alternative to a statistical model. The DTW handles variations in time and position which allows a dynamic variation in the temporal pace of gesture. Gestures are compared to a codebook of reference gestures by finding the best-case dynamic temporal distortion to compare two gestures. Dynamic temporal distortion consists an approach using a uniform time scaling that assumes the entire gesture is performed with a single speed factor. This avoids requiring a reference gesture for every possible timing variation. Variations in the positional path are handled by providing a representative set of gesture variations in the codebook. This set is sufficient enough to classify gestures using unsupervised clustering algorithms such as a single nearest neighbor or a k-nearest neighbor. An optimal choice of dynamic time warping and minimum error is based on a euclidean distance metric.

Algorithm2. Dynamic Time Warping(O, M)


Input: $O = [(x_1, y_1), \dots, (x_n, y_n)]$ (positions for the gesture)
 $M = [(x'_1, y'_1), \dots, (x'_n, y'_n)]$ (model positions for the gesture)
Output: d (Warped Distance),
 $d(0,0) = \text{distance}(M(1), O(1))$
for i:=1 **to** len(O) **do**
 $d(i, 1) = d(i - 1, 1) + \text{distance}(M(1), O(i))$
end for
for j: =1 **to** len(M) **do**
 $d(1, j) = d(1, j - 1) + \text{distance}(M(j), O(1))$
end for
for j: =1 **to** len(M) **do**
for i:=1 **to** len(G) **do**
 $d(1, j) = \min_{i,j} d[d(i-1, j-1), d(i-1, j), d(i, j-1)] + \text{distance}(M(j), O(i))$
end for
end for
return $d(n, m)$
Complexity: $O(n \times m)$

5. System Evolution

The goal of this system is to provide a highly accurate gesture recognition so that users can adopt to changes in the position and orientation of the sensor. The sensor array size and rotation according to the user hand. The HMM and DTW model would ideally need to be retrained on the gestures for every rotation. To overcome this problem a system is designed that augments the capacitive sensor array with wrist worn accelerometer band, this band could be replaced by smart watch.

5.1 Experiment Setup

We tested this system on five subjects, who has C6 spinal cord injury. For each session subject were asked to perform the five gestures, during the experiment ten of each type of gesture at different rotation of degrees of the sensor array, hence a total 150 gestures performed per subject.






Figure 4: (a) Confusion matrix illustrating the accuracy of recognizing the five gestures performed by all the subjects when our system used Dynamic Time Warping. (b) Confusion matrix illustrating the accuracy of recognizing the gestures performed by the subjects when our system used Hidden Markov Model. For both figures, the confusion matrix is calculated using $150 \times 5 = 750$ gestures. Our system has an average accuracy of 99 and 97 percent for the DTW and HMM models respectively.

The figure 4 illustrate the classification of accuracy of gesture recognition algorithm. The figure generated using data from all gestures performed by five subjects when sensor array rotated different degrees with respect to hand. The average accuracy is 99 and 97 percent of DTW and HMM respectively. If there is a large variation in the way gestures performed then HMM model would perform best while the DTW would perform poorly.




Figure 5: Accuracy per subject for five gestures when the classification algorithm uses a DTW model and an HMM classification.

The figure 5 illustrates the accuracy of the DTW and HMM based gesture recognition per subject. The height bar shows the accuracy across five gestures while the error bar shows the standard deviation in accuracy across gestures per subject. The focus of the experiment on the subjects who has C6 spinal cord injury. The subject perform some set of gestures and it shows that the accuracy of DTW algorithm on the gestures is 99 percent, Hence we can say that this system works very well even though the subject use limited mobility.

6. Conclusion

In this paper we proposed a system which uses the textile capacitive sensor arrays to recognize gestures for the users who are physically disable and unable to do their routine, this system can provide gesture recognition with limited mobility. The textile sensors are flexible and can built into clothing or daily use items such as pillow covers, bed sheets, wheelchair pads etc. An adaptive and personalized signal processing system can convert capacitance data from sensor array to alphanumeric gestures. The gestures can use then to control the appliances in the home. By using the wrist-worn accelerometer and the sensor array, the gesture recognition system correct the rotation of the sensor array with the hand with minimal intervention. This system has evaluated by using the HMM and DTW classification. The DTW is more computationally efficient option.

References

- [1] Alexander Nelson, Student Member, IEEE, Gurashish Singh, Member, IEEE, Ryan Robucci, Member, IEEE, Chintan Patel, Member, IEEE, and Nilanjana Banerjee "Adaptive and Personalized Gesture Recognition Using Textile Capacitive Sensor Arrays".
- [2] A. Lang, W. Ruland-Brown, and M. M. Wald, "The epidemiology and impact of traumatic brain injury: A brief overview," *J. Head Trauma Rehabil.*, vol. 5, pp. 375–378, 2006.
- [3] National spinal cord injury statistical center, facts and figures at a glance, Birmingham, AL: University of Alabama at Birmingham, 2015.
- [4] G. Schmeissner and W. Seamone, "An assistive equipment controller for quadriplegics," *Johns Hopkins Med. J.*, 143, no. 3, pp. 84–88, 1979.
- [5] S. Strassman, J. D. Harviel, J. H. Jaffin, and M. H. Jordan, "Lightning strike to the head: Case report," *J. Trauma Acute Care Surgery*, vol. 36, no. 1, pp. 113–115, 1994.
- [6] M. S.Workinger and R.Netsell, "Restoration of intelligible speech 13 years post-head injury," *Brain Injury*, vol. 6, no. 2, pp. 183–187, 1992.
- [7] H. Junker, O. Amft, P. Lukowicz, and G. Tröster, "Gesture spotting with body-worn inertial sensors to detect user activities," *Pattern Recog.*, vol. 41, no. 6, pp. 2010–2024, 2008.
- [8] A. Y. Benbasat and J. A. Paradiso, "An inertial measurement framework for gesture recognition and applications," in *Proc. Int. Gesture Workshop Gesture Sign Language Human-Comput. Interaction*, 2002, pp. 9–20.
- [9] S.-J. Cho, J. K. Oh, W.-C. Bang, W. Chang., E. Choi, Y. Jing, J. Cho, and D. Y. Kim, "Magic wand: A hand-drawn gesture input device in 3-d space with inertial sensors," in *Proc. IEEE 9th Int. Workshop Frontiers Handwriting Recog.*, 2004, pp. 106–111.
- [10] Y. Wu and T. S. Huang, "Vision-based gesture recognition: A review," in *Proc. Int. GestureWorkshop Gesture-Based Commun. Human-Comput. Interaction*, 1999, pp. 103–115. [11] Y. Xia, Z.-M. Yao, X.-J. Yang, S.-Q. Xu, X. Zhou, and Y.-N. Sun,
- [11] "A footprint tracking method for footprint tracking method for gait analysis," *Biomed. Eng.: Appl.*, Basis

- Commun., vol. 26, no. 1, 2014, Doi: 10.4015/S1016237214500112.
- [12] S. K. Fager and J. M. Burnfield, "Patients' experiences with technology during inpatient rehabilitation: Opportunities to support independence and therapeutic engagement," *Disability Rehabil.: Assistive Technol.*, vol. 9, no. 2, pp. 121–127, 2013.
- [13] J. Cheng, O. Amft, and P. Lukowicz, "Active capacitive sensing: Exploring a new wearable sensing modality for activity recognition," in *Proc. 8th Int. Conf. Pervasive Comput.*, 2010, pp. 319–336.
- [14] H. K. Ouh, J. Lee, S. Han, H. Kim, I. Yoon, and S. Hong, "A programmable mutual capacitance sensing circuit for a large-sized touch panel," in *Proc. IEEE Int. Symp. Circuits Syst.*, 2012, pp. 1395–1398.
- [15] Y. Piao, X. Xiao, X. Wang, Q. Zhou, K. Fan, and P. Liu, "Conditioning circuit for capacitive position sensor with nanoscale precision based on AC excitation principle," in *Proc. IEEE Int. Conf. Electro/Inf. Technol.*, 2011, pp. 1–6.
- [16] P.-H. Lo, C. Hong, S.-H. Tseng, J.-H. Yeh, and W. Fang, "Implementation of vertical-integrated dual mode inductivecapacitive proximity sensor," in *Proc. IEEE 25th Int. Conf. Micro Electro Mech. Syst.*, pp. 640–643.
- [17] H. Jiang, Z. Han, P. Scucces, S. Robidoux, and Y. Sun, "Voiceactivated environmental control system for persons with disabilities," in *Proc. Bioeng. Conf.*, 2010, pp. 167–168.
- [18] S. B. Kang, "Hands-free interface to a virtual reality environment using head tracking," *U.S. Patent 6 006 210*, Dec. 28, 1999.
- [19] A. Nelson, J. Schmandt, P. Shyamkrishna, W. Jenkins, D. Lachut, N. Banerjee, S. Rollins, J. Parkerson, and V. Varadan, "Wearable multi-sensor gesture recognition for paralysis patients," in *Proc. IEEE Sensors*, 2013, pp. 1–4.
- [20] A. D. Wilson and A. F. Bobick, "Parametric hidden Markov models for gesture recognition," *IEEE Trans. Pattern Anal. Mach. Intell.*, vol. 21, no. 9, pp. 884–890, Sep. 1999.
- [21] G. Fang, W. Gao, and D. Zhao, "Large vocabulary sign language recognition based on hierarchical decision trees," in *Proc. 5th Int. Conf. Multimedia Interf.*, 2003, pp. 125–131.
- [22] P. Domingos and M. J. Elzazzani, "On the optimality of the simple bayesian classifier under zero-one loss," *Mach. Learn.*, vol. 9, no. 2-3, pp. 103–121, 1997.
- [23] T. Starner, J. Makhoul, R. Schwartz, and G. Chou, "On-line cursive handwriting recognition using speech recognition methods," in *Proc. IEEE Int. Conf. Acoust., Speech, Signal Process.*, 1994, pp. V/125–V/128.
- [24] Microchip, Single-Zone Tracking and Gesture Controller Data Sheet, P. 100, C, Nov. 2013.
- [25] J. Hershberger and J. Snoeyink, "An O(nlogn) implementation of the douglas-peucker algorithm for line simplification," in *Proc. 10th Annu. Symp. Comput. Geometry*, 1994, pp. 383–384.
- [26] J. O. Wobbrock, B. A. Myers, and J. A. Kembel, "Edgewrite: a stylus-based text entry method designed for high accuracy and stability of motion," in *Proc. 16th Annu. ACM Symp. User Interface Softw. Technol.*, 2003, pp. 61–70.